




This response card should be provided to the recipient of the package the indicator is placed in; enabling them to obtain proper reading of the indicator. Please call for additional usage information.

**3M**

HCL #8207-01

**MonitorMark™**  
**10 L Time/Temperature Tag**  
 Response Temperature 50°F (10°C)








This response card should be provided to the recipient of the package the indicator is placed in; enabling them to obtain proper reading of the indicator. Please call for additional usage information.

**3M**

HCL #8207-01

**MonitorMark™**  
**10 L Time/Temperature Tag**  
 Response Temperature 50°F (10°C)



The above illustration shows how to use the runout indexes on the other side of this card to estimate the time a particular activated tag has been exposed above its Response Temperature

From the readings on the two index scales and some general knowledge about the exposure conditions, one can estimate time/temperature exposures



**3M**

Identification and Converter Systems Division

3M Center, Building 220-7W-03  
 St Paul, MN 55144-1000

REORDER FROM  
**HEALTH CARE LOGISTICS**  
 CALL FREE 1-800-848-1633  
 FAX FREE 1-800-447-2923

HCL #8207-01  
 Made in U.S.A.



The above illustration shows how to use the runout indexes on the other side of this card to estimate the time a particular activated tag has been exposed above its Response Temperature

From the readings on the two index scales and some general knowledge about the exposure conditions, one can estimate time/temperature exposures

**3M**

Identification and Converter Systems Division

3M Center, Building 220-7W-03  
 St Paul, MN 55144-1000

REORDER FROM  
**HEALTH CARE LOGISTICS**  
 CALL FREE 1-800-848-1633  
 FAX FREE 1-800-447-2923

HCL #8207-01  
 Made in U.S.A.